



EC-type examination certificate

Number **T10162** revision 1
Project number 9200235
Page 1 of 1

Issued by NMI Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands

In accordance with The "Metrologiewet" (Stb. 2006, 137) as Dutch implementation of Directive 2004/22/EC on measuring instruments (MID).

Manufacturer Precia Molen
Franse Akker 1
4824 AL Breda
The Netherlands

In respect of A model of a **Discontinuous totalizing automatic weighing instrument**
Type: ABS-...-(SCS if applicable)

Characteristics

Accuracy class (OIML R107):	0.2, 0.5, 1 and 2
Climatic environment:	-10 °C / +40 °C 85% humidity at 40 °C non-condensing
Electromagnetic environment class:	E2
Mechanical environment class:	not applicable

In the description number T10162 revision 1 further characteristics are described.

Valid until 17 September 2018

Description and documentation The instrument is described in the description number T10162 revision 1 and documented in the documentation folder T10162-2, appertaining to this EC-type examination certificate.

Remarks This revision EC type-approval certificate replaces the earlier versions, including its documentation folder.

The Notified Body no. 0122
NMI Certin, 3 June 2009


C. Oosterman
Head Certification Board

NMI Certin BV
Hugo de Grootplein 1
3314 EG Dordrecht
PO Box 394
3300 AJ Dordrecht, NL
T +31 78 6332332
F +31 78 6332309
certin@nmi.nl
www.nmi.nl

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMI (see "Regulation objection and appeal against decisions of NMI")

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

Reproduction of the complete document is permitted.

